


S.no.	Date	Department/Place	Complaint	Type of Problem (Mechanical/Electrical/ Plumbing/Any other)	Signature of the Complainee	Work Done	Status of the workdone	Workdone By	Remarks if any	Referred To (further referred if not corrected in college premises)	Follow Up	Signature of the supervisor
161	20/4/22	OP	Lab - ceiling fan slow movement	electrical	Dr. Nalime	Capacitor changed	Completed on 25/4/22	Bakshush	work done		3/5/22	John
162	22/4/22	Redo (A)	Suction is not working	Mechanical	Dr. Gargi	Pipeline clean & repaired	Completed on 29/4/22	Rajeev	effective		3/5/22	John
163	23/4/22	Redo (A)	Perforation in pipeline (ultrasonic cleaner)	Mechanical	Dr. Heerjit	Pipe changed	Completed 26/4/22	Ma-deep	work completed		6/5/22	John
164	24/4/22	OP	OP area - ceiling fan not working	electrical	Dr. Nalime	Repaired	Completed on 30/4/22	Bakshush	done		10/5/22	John
165	24/4/22	OS	UV chamber - tube light not working (disturbance from)	Electrical	Dr. Akash	Tube light changed	Completed on 26/4/22	Raj Patel	effective		6/5/22	John


S.no.	Date	Department/Place	Complaint	Type of Problem (Mechanical/Electrical/ Plumbing/Any other)	Signature of the Complainee
1016	21/12/21	MC	Water not working (Bulbabin)	Mechanical	Dr. Kunal
1017	22/12/21	Redo (H)	Dental chair S not working	Mechanical	Dr. Saumya
1018	23/12/21	Redo (H)	Pressure leakage from dental chair	Mechanical	Dr. Prabak
1019	23/12/21	ortho	Chair not working	Mechanical	Dr. SSI
1020	27/12/21	Redo (H)	Suction not working	Mechanical	Dr. Anand

Work Done	Status of the workdone	Workdone By	Remarks if any	Referred To (further referred if not corrected in college premises)	Follow Up	Signature of the supervisor
Solenoid valve changed	Completed on 23/12/21	Rajeev	done		31/122	Dr. Kunal
Pipe replaced	Completed on 23/12/21	Mandeep	effective		31/122	Dr. Saumya
Connection tightened	Completed on 22/12/21	Rajeev	Completed		31/122	Dr. Prabak
Spot control fixed	Completed on 24/12/21	Mandeep	effective		4/1/22	Dr. SSI
Valve changed	Completed on 29/12/21	Mandeep	done		31/122	Dr. Anand




S.no.	Date	Department/Place	Complaint	Type of Problem (Mechanical/Electrical/ Plumbing/Any other)	Signature of the Complainee
491	16/12/21	Lab 00	Two ultrasonic scalers not working.	Mechanical	Niti
492	17/12/21	Lab 10	Heckelaw not working	electrical	Dr Soni
493	20/12/21	Gas (U4)	Flue light not working	electrical	Dr Sagar
494	20/12/21	Protho (H)	Micrometer not working (analog lab)	Mechanical	Dr Kadli
495	21/12/21	Protho (U4)	Fan not working	electrical	Dr Viveet

Work Done	Status of the workdone	Workdone By	Remarks if any	Referred To (further referred if not corrected in college premises)	Follow Up	Signature of the supervisor
Replace new tip	Completed on 20/12/21	Mandeep	done		30/12/21	Dr Soni
Power switch replaced	Completed on 18/12/21	Bakshish	Completed		28/12/21	Dr Soni
Replaced with new	Completed on 22/12/21	Bakshish	work done		21/1/22	Dr Soni
Control box connected	Completed on 21/12/21	Mandeep	effective		21/1/22	Dr Soni
Fan replaced	Completed on 25/12/21	Bakshish	Completed		31/1/22	Dr Soni




S.no.	Date	Department/Place	Complaint	Type of Problem (Mechanical/Electrical/ Plumbing/Any other)	Signature of the Complainee
276	16/12/20	Pho	Andalase not working	Mechanical	Dr. S. S. K. K.
277	17/12/20	Co (U)	Skullgation room racks bottle	any other	Dr. S. S. K. K.
278	17/12/20	Pedo	Chair 4 - not working	electrical	Dr. S. S. K. K.
279	18/12/20	Pedro	Print of the page is not clear	any other	Nitin
280	18/12/20	OP	Repl library - Tubelyst not working	electrical	Dr. S. S. K. K.


Work Done	Status of the workdone	Workdone By	Remarks if any	Referred To (further referred if not corrected in college premises)	Follow Up	Signature of the supervisor
Repaired	Completed on 17/12/20	Hardeep	Completed		28/12/20	
Har racks are fixed	Completed on 18/12/20	Shalini	done		28/12/20	
Main light source wire changed	Completed on 19/12/20	Dakshin	effective		29/12/20	
Refill the cartridge	Completed on 19/12/20	Arjun	work done		29/12/20	
Corrected	Completed on 24/12/20	Dakshin	Completed		2/1/21	




S.no.	Date	Department/Place	Complaint	Type of Problem (Mechanical/Electrical/ Plumbing/Any other)	Signature of the Complainer	Work Done	Status of the workdone	Workdone By	Remarks if any	Referred To (further referred if not corrected in college premises)	Follow Up	Signature of the supervisor
286	23/12/20	Pho	Tubelight of Sterilization Room not working	electrical	Dr Soni	Switch replaced	Completed on 23/12/20	Bakshesh	Completed		2/1/21	Dr Soni
287	22/12/20	Recs	Chair no 2 is not moving up & down	mechanical	Dr Nitin	Repair P.B & motor of chair	Completed on 22/12/20	Mandeep	done		2/1/21	Dr Nitin
288	23/12/20	Pho (H)	Switch broken (dept. library)	electrical	Dr Tejansh	Switch replaced	Completed on 23/12/20	Bakshesh	effective		4/1/21	Dr Tejansh
289	24/12/20	Gen (Ph)	Endometer is not working	Any other	Dr Vidhi	Endometer repaired	Completed on 30/12/20	Mandeep	done		10/1/21	Dr Vidhi
300	26/12/20	Recs	Chair no 4, Scaler not working	mechanical	Piyush	Replace new Scaler hand piece	Completed on 26/12/20	Mandeep	Completed		5/1/21	Dr Piyush



S.no.	Date	Department/Place	Complaint	Type of Problem (Mechanical/Electrical/ Plumbing/Any other)	Signature of the Complainee	Work Done	Status of the workdone	Workdone By	Remarks if any	Referred To (further referred if not corrected in college premises)	Follow Up	Signature of the supervisor
436	12/12/19	OS	fan not working	electrical	Dr. Arin	Switches replaced	Completed on 20/12/19	Cap 2	Completed		30/12/19	
437	17/12/19	PHD	Tube light not working	electrical	Dr. Sameer	Tube light replaced	Completed on 18/12/19	Balishish	work done		28/12/19	
438	18/12/19	CMR	M/C is not working properly in the O/D area	Mechanical	Dr. Divya	M/C changed	Completed on 18/12/19	Balishish	done		28/12/19	
439	18/12/19	CMR	Door lock has been broken in reader room	Any other	Dr. Divya	lock changed	Completed on 18/12/19	Skateed	done		28/12/19	
440	18/12/19	PHD	Gas control - gas control is not working properly	Mechanical	Chander	Replace new gas control	Completed on 19/12/19	Ranjit	effective		30/12/19	



S.no.	Date	Department/Place	Complaint	Type of Problem (Mechanical/Electrical/ Plumbing/Any other)	Signature of the Complainee	Work Done	Status of the workdone	Workdone By	Remarks if any	Referred To (further referred if not corrected in college premises)	Follow Up	Signature of the supervisor
441	21/12/19	ortho	chair light not working	electrical	Dr. Singh	bulb replaced	Completed on 24/12/19	mandap	done		3/1/20	Dr. Singh
442	23/12/19	Restro (UG)	Turner not working	Mechanical	Dr. Buehli	Disc changed	Completed on 24/12/19	Rajeev	effective	Sent to company for repair	3/1/20	Dr. Singh
443	23/12/19	OS	dental chair bulb not working	electrical	Dr. Kamm	bulb replaced	Completed on 27/12/19	Rajeev	work Completed		6/1/20	Dr. Singh
444	23/12/19	Labo	Stater no 3 not working in chair no 6	Mechanical	Dr. Kishabh	water pipe changed	Completed on 24/12/19	mandap	done		3/1/20	Dr. Singh
445	23/12/19	Gen (UG)	Dental Unit No-12 - chair knobs broken	anyother	Dr. Suresh	Chair knobs has been replaced	Completed on 28/12/19	Rajeev	work done		6/1/20	Dr. Singh



451 Size IV Industrial Area, Sahibabad
Gurgaon, Haryana 201018
MAINTENANCE REGISTER

Indraprastha Dental College & Hospital
451, Size IV Industrial Area, Sahibabad
Gurgaon, Haryana 201018
MAINTENANCE REGISTER


S.no.	Date	Department/Place	Complaint	Type of Problem (Mechanical/Electrical/ Plumbing/Any other)	Signature of Complainer	Work Done	Status of the workdone	Workdone By	Remarks if any	Referred To (further referred if not corrected in college premises)	Follow Up	Signature of the supervisor
456	23/12/18	Recor	Therapy ultrasonic not working	Electrical	Dr. Nisha	Repaired	Completed 22/12/18	Rekhat	Done		24/12/18	John
457	23/12/18	OS	Hand tool was staying from chair (dentist chair)	Any other	Dr. Nisha	Food that replaced	Completed 24/12/18	Rekhat	Done		31/12/18	John
458	24/12/18	Recor	ultrasonic cleaner not working	Mechanical	Dr. Nisha	Repaired PCB	Completed 31/12/18	Hardy	effective		31/1/19	John
459	26/12/18	Other	chair not working	Mechanical	Dr. Nisha	dent wire replaced	Completed 24/12/18	Hardy	stand broken		2/1/19	John
460	26/12/18	Pink	generator not working (recharge)	Electrical	Dr. Nisha	Electrical fault checked	27/12/18	Rajpal	Done		4/1/19	John



Inderprastha Dental College & Hospital
Plot No. 14 Industrial Area, Sahibabad
(Nearby: Old Post Office) 201015
MAINTENANCE REGISTER

S.no.	Date	Department/Place	Complaint	Type of Problem (Mechanical/Electrical/ Plumbing/Any other)	Signature the Complainant	Work Done	Status of the workdone	Workdone By	Remarks if any	Referred To (further referred if not corrected in college premises)	Follow Up	Signature of the supervisor
461	27/12/18	OS	Water leakage from bottle (dental chair)	Any other	Dr. Garg	Water replaced	Completed on 27/12/18	Ranjit	Done		27/12/18	
462	28/12/18	Cards	1/2 water leakage (computer)	Plumbing	Dr. Singh	Water leakage fixed	Completed on 27/12/18	Ramendra	Done		27/12/18	
463	28/12/18	Radio	Radio image not clear	Mechanical	Dr. Jaiswal	Radio is repaired	Completed on 29/12/18	Mohd. Arif	Radio stand		31/12/18	
464	28/12/18	Restroom	Broken switch (wash up room)	Electrical	Dr. Singh	Switch replaced	Completed on 29/12/18	Sigra	Done		27/12/18	
465	28/12/18	astro	Projector not working	Electrical	Dr. Singh	Projector replaced	Completed on 29/12/18	F-11	Effective		31/12/18	

Inderprastha Dental College & Hospital
Plot No. 14 Industrial Area, Sahibabad
(Nearby: Old Post Office) 201015
MAINTENANCE REGISTER



S.no.	Date	Department/Place	Complaint	Type of Problem (Mechanical/Electrical/ Plumbing/Any other)	Signature the Work Done Complain	Status of the workdone	Workdone By	Remarks if any	Referred To (further referred if not corrected in college premises)	Follow Up	Signature of the supervisor
45	18/12/17	Caf	Chair in 2nd room not working	Mechanical	Dr. Nigra	Completed on 19/12/17	Handley	Done		20/12/17	
46	19/12/17	Lab	Drill machine not working	Mechanical	Dr. Nigra	Completed on 20/12/17	Handley	Done		20/12/17	
47	20/12/17	Lab	Drill machine not working	Mechanical	Dr. Nigra	Completed on 21/12/17	Handley	Done		21/12/17	
48	21/12/17	Lab	Drill machine not working	Mechanical	Dr. Nigra	Completed on 22/12/17	Handley	Done		22/12/17	
49	21/12/17	Lab	Drill machine not working	Mechanical	Dr. Nigra	Completed on 23/12/17	Handley	Done		23/12/17	



S.no.	Date	Department/Place	Complaint	Type of Problem (Mechanical/Electrical/ Plumbing/Any other)	Signature the Work Done Complain	Status of the workdone	Workdone By	Remarks if any	Referred To (further referred if not corrected in college premises)	Follow Up	Signature of the supervisor
40	29/12/17	PHD	Solar (water) satellite not working	Mechanical	Dr. Pankaj	Regulator changed	23/12/17	Mansley	Done	30/12/17	J. Khan
41	24/12/17	PHD	Cham of alarm switch not working	Mechanical	Dr. Pankaj	Switch replaced	26/12/17	Pankaj	Done	31/12/17	J. Khan
42	26/12/17	OS	Water pump not working	Electrical	Dr. Pankaj	Replaced	Completed on 29/12/17	syim	Effected	31/12/17	J. Khan
43	27/12/17	PHD	Cracked handle (Casting lost)	Mechanical	Dr. Pankaj	Handle replaced	Completed on 28/12/17	Mansley	Solved	4/1/18	J. Khan
44	28/12/17	PHD	Dental Station 15 Cham broken	Mechanical	Dr. Pankaj	Cham Knives Replaced	Completed on 30/12/17	Pankaj	Solved	4/1/18	J. Khan



2017

S.no.	Date	Department/Place	Complaint	Type of Problem (Mechanical/Electrical/ Plumbing/Any other)	Signature of the Complainee
1	3/1/17	Oral	Remixer not working due to damaged plug	Electrical	D. Jaisankar
2	3/1/17	Oral	Bitler not working	Mechanical	D. Siddharth
3	3/1/17	Oral	Suction not working	Any other	D. Siddharth
4	3/1/17	Oral	Photopys not working (Kumar Jaisankar)	Electrical	D. Kishi
5	3/1/17	General	Chair no 12 not working	Any other	L. Naga

Work Done	Status of the workdone	Workdone By	Remarks if any	Referred To (further referred if not corrected in college premises)	Follow Up	Signature of the supervisor
Remixer plug changed	Completed on 5/1/17	Sajam	Done		12/1/17	J. Kumar
Bitler replaced	Completed on 4/1/17	Manday	Done		12/1/17	J. Kumar
HP holder replaced	Completed on 5/1/17	Parjit	Effective		12/1/17	J. Kumar
Repaired	Completed on 3/1/17	Sajam	Effective		11/1/17	J. Kumar
Chair repair	Completed on 9/1/17	Parjit	Solved		10/1/17	J. Kumar

